

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002578  
7:34 17/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:56  
PRESSURIZING 3:03  
STERILIZING 4:09  
136.1 C 221kPa 4:09  
136.4 C 225kPa 4:39  
136.3 C 224kPa 5:09  
135.8 C 221kPa 5:39  
136.3 C 223kPa 6:09  
136.2 C 221kPa 6:39  
136.2 C 221kPa 7:09  
136.4 C 223kPa 7:39  
136.4 C 223kPa 7:39  
VENTING 7:39  
AIR DRYING 8:31  
CYCLE COMPLETE 8:41  
-----

Digital Signature #  
4118224C315E3E7D  
-----