

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2555

Name: RITTER 1  
Sterilizer ID: V2178198  
11/14/2022 04:52 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 206.3 F 0.0  
2:00 195.2 F 4.9  
4:00 209.3 F 10.2  
6:00 220.5 F 5.5  
8:00 236.0 F 11.7  
10:00 236.1 F 9.2  
12:00 238.7 F 9.0  
14:00 249.8 F 14.9  
16:00 262.5 F 22.2  
18:00 272.1 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.4 F 28.8  
2:00 271.4 F 28.1  
3:00 272.5 F 29.0  
4:00 271.9 F 28.3  
  
Min 271.3 F 28.0  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.2 F 0.0  
5:00 221.3 F 0.0  
10:00 241.8 F 0.0  
15:00 231.6 F 0.0  
20:00 228.3 F 0.0  
25:00 224.6 F 0.0  
30:00 223.6 F 0.0  
Drying Complete

FILLING: 2:10  
HEATING: 18:11  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:57:34  
CYCLE COMPLETE

11/14/2022 05:49 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_