
Midmark M11 - v1.0.5
Total Cycles: 2929

Name: RITTER2
Sterilizer ID: V2313228
09/26/2023 11:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.0 F 0.0
2:00 174.7 F 2.9
4:00 191.7 F 7.8
6:00 207.3 F 13.1
8:00 223.2 F 8.2
10:00 236.0 F 13.5
12:00 238.4 F 11.2
14:00 237.8 F 9.8
16:00 250.9 F 16.2
18:00 262.2 F 22.6
20:00 272.7 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.4 F 29.5
2:00 272.5 F 29.4
3:00 272.3 F 29.4
4:00 272.4 F 29.4

Min 270.7 F 28.6
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.7 F 0.0
5:00 211.6 F 0.0
10:00 209.4 F 0.0
15:00 218.1 F 0.0
20:00 216.3 F 0.0
25:00 215.9 F 0.0
30:00 211.8 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 20:18
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 00:59:32
CYCLE COMPLETE

09/26/2023 12:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____