
Midmark M11 - v1.0.5
Total Cycles: 3172

Name: RITTER2
Sterilizer ID: V2313228
11/30/2023 12:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.0 F 0.0
2:00 110.2 F 0.1
4:00 117.0 F 1.3
6:00 143.2 F 3.8
8:00 169.1 F 7.1
10:00 190.6 F 10.9
12:00 211.8 F 4.7
14:00 216.5 F 8.0
16:00 231.1 F 12.3
18:00 232.1 F 8.9
20:00 239.9 F 11.3
22:00 243.6 F 12.9
24:00 255.1 F 18.9
26:00 265.7 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.4
1:00 272.9 F 30.2
2:00 271.8 F 29.5
3:00 271.5 F 29.7
4:00 272.6 F 30.1

Min 271.2 F 29.2
Max 273.1 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.7 F 0.0
5:00 223.3 F 0.0
10:00 212.7 F 0.0
15:00 194.7 F 0.0
20:00 191.3 F 0.0
25:00 188.4 F 0.0
30:00 206.9 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 27:50
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:06:13
CYCLE COMPLETE

11/30/2023 01:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____