

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002534  
8:41 24/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:54  
PRESSURIZING 2:56  
STERILIZING 4:00  
136.0 C 220kPa 4:00  
136.4 C 225kPa 4:30  
136.4 C 223kPa 5:00  
136.0 C 221kPa 5:30  
136.3 C 223kPa 6:00  
136.1 C 220kPa 6:30  
136.4 C 224kPa 7:00  
136.2 C 223kPa 7:30  
136.2 C 223kPa 7:30  
VENTING 7:30  
AIR DRYING 8:21  
CYCLE COMPLETE 9:06  
-----

Digital Signature #  
0B3C7B2146CD3708  
-----