
Midmark M11 - v1.0.0
Total Cycles: 4716

Name: RITTER 1
Sterilizer ID: V2178198
07/11/2024 11:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.2 F 0.0
2:00 144.1 F 2.3
4:00 161.4 F 5.5
6:00 184.6 F 9.1
8:00 202.8 F 12.9
10:00 218.6 F 4.5
12:00 228.9 F 8.4
14:00 238.7 F 12.8
16:00 236.7 F 9.2
18:00 240.2 F 10.5
20:00 246.1 F 13.1
22:00 257.4 F 19.1
24:00 267.7 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.1
1:00 272.5 F 28.9
2:00 271.7 F 28.3
3:00 271.7 F 28.7
4:00 272.7 F 29.1

Min 271.0 F 28.0
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.1 F 0.0
5:00 225.7 F 0.0
10:00 225.4 F 0.0
15:00 229.7 F 0.0
20:00 228.8 F 0.0
25:00 226.3 F 0.0
30:00 235.3 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 25:29
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:04:17
CYCLE COMPLETE

07/11/2024 12:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____