
Midmark M11 - v1.0.0
Total Cycles: 3095

Name: RITTER 1
Sterilizer ID: V2178198
04/11/2023 04:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.3 F 0.0
2:00 178.4 F 3.4
4:00 196.7 F 7.6
6:00 215.4 F 12.9
8:00 224.7 F 7.9
10:00 238.1 F 13.2
12:00 237.6 F 10.4
14:00 235.0 F 8.4
16:00 249.6 F 15.4
18:00 261.8 F 22.2
20:00 272.6 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.6 F 29.3
2:00 271.5 F 28.8
3:00 272.6 F 29.4
4:00 271.6 F 28.6

Min 271.4 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.3 F 0.0
5:00 229.7 F 0.0
10:00 238.9 F 0.0
15:00 236.0 F 0.0
20:00 227.5 F 0.0
25:00 221.5 F 0.0
30:00 219.7 F 0.0
Drying Complete

FILLING: 1:44
HEATING: 20:25
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:24
CYCLE COMPLETE

04/11/2023 05:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____