
Midmark M11 - v1.0.5
Total Cycles: 844

Name: RITTER2
Sterilizer ID: V2313228
02/28/2022 12:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.2 F 0.0
2:00 181.1 F 4.0
4:00 206.3 F 9.7
6:00 221.3 F 6.7
8:00 234.5 F 11.9
10:00 237.5 F 10.3
12:00 237.9 F 9.7
14:00 252.8 F 17.2
16:00 265.3 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.7 F 29.5
2:00 271.4 F 28.7
3:00 272.3 F 29.7
4:00 272.5 F 29.2

Min 271.2 F 28.7
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.2 F 0.0
5:00 214.8 F 0.0
10:00 215.0 F 0.0
15:00 217.8 F 0.0
20:00 220.5 F 0.0
25:00 218.9 F 0.0
30:00 219.9 F 0.0
Drying Complete

FILLING: 2:22
HEATING: 17:42
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:57:10
CYCLE COMPLETE

02/28/2022 01:33 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____