

-----  
Midmark M11 - v1.0.5  
Total Cycles: 588

Name: RITTER2  
Sterilizer ID: V2313228  
11/30/2021 12:10 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.3 F 0.0  
2:00 161.6 F 3.0  
4:00 182.3 F 7.4  
6:00 206.3 F 13.0  
8:00 221.0 F 7.2  
10:00 234.2 F 12.6  
12:00 237.2 F 10.4  
14:00 235.8 F 9.0  
16:00 250.7 F 16.1  
18:00 262.8 F 23.2  
20:00 272.3 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.0  
1:00 272.5 F 29.5  
2:00 271.9 F 29.5  
3:00 272.3 F 29.4  
4:00 271.8 F 29.4

Min 271.5 F 28.9  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.0 F 0.0  
5:00 226.6 F 0.0  
10:00 228.9 F 0.0  
15:00 227.2 F 0.0  
20:00 222.4 F 0.0  
25:00 220.1 F 0.0  
30:00 216.3 F 0.0  
Drying Complete

FILLING: 2:05  
HEATING: 20:10  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:59:23  
CYCLE COMPLETE

11/30/2021 01:26 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02