
Midmark M11 - v1.0.5
Total Cycles: 1445

Name: RITTER2
Sterilizer ID: V2313228
09/02/2022 07:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.5 F 0.0
2:00 75.4 F 0.0
4:00 83.2 F 0.0
6:00 106.7 F 0.0
8:00 140.1 F 1.3
10:00 168.1 F 3.1
12:00 192.7 F 5.5
14:00 207.6 F 8.8
16:00 221.3 F 12.5
18:00 225.0 F 7.3
20:00 235.7 F 11.9
22:00 235.7 F 9.0
24:00 240.9 F 10.5
26:00 246.0 F 13.4
28:00 257.2 F 19.5
30:00 267.5 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.3 F 29.3
2:00 271.5 F 28.9
3:00 272.3 F 29.2
4:00 271.5 F 29.0

Min 271.3 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.2 F 0.0
5:00 201.8 F 0.0
10:00 215.0 F 0.0
15:00 215.8 F 0.0
20:00 218.1 F 0.0
25:00 214.9 F 0.0
30:00 219.8 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:28
STERILIZING: 4:00
VENTING: 3:00
DRYING: 30:00
TOTAL CYCLE: 01:09:18
CYCLE COMPLETE

09/02/2022 08:25 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____