

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3927

Name: RITTER2  
Sterilizer ID: V2313228  
09/11/2024 09:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.1 F 0.0  
2:00 70.4 F 0.0  
4:00 74.3 F 0.0  
6:00 104.1 F 0.0  
8:00 147.4 F 0.1  
10:00 182.4 F 0.2  
12:00 207.2 F 1.2  
14:00 210.8 F 4.0  
16:00 216.5 F 6.9  
18:00 226.1 F 10.1  
20:00 233.8 F 11.7  
22:00 235.1 F 9.3  
24:00 243.5 F 12.1  
26:00 242.9 F 11.9  
28:00 240.4 F 10.6  
30:00 252.2 F 16.5  
32:00 262.5 F 22.5  
34:00 271.9 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 28.5  
1:00 272.3 F 29.2  
2:00 272.0 F 28.5  
3:00 272.3 F 29.2  
4:00 272.1 F 28.7  
  
Min 271.4 F 28.3  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.4 F 0.0  
5:00 233.1 F 0.0  
10:00 218.3 F 0.0  
15:00 176.6 F 0.0  
20:00 173.4 F 0.0  
25:00 172.2 F 0.0  
30:00 170.8 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 34:25  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 01:12:48  
CYCLE COMPLETE

09/11/2024 10:19 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_