
Midmark M11 - v1.0.0
Total Cycles: 1886

Name: RITTER 1
Sterilizer ID: V2178198
04/28/2022 01:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.8 F 0.0
2:00 184.5 F 3.1
4:00 184.0 F 7.2
6:00 209.6 F 12.3
8:00 221.9 F 6.7
10:00 236.2 F 11.9
12:00 234.7 F 8.7
14:00 241.0 F 10.1
16:00 246.6 F 13.5
18:00 259.2 F 20.4
20:00 270.6 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.5 F 29.0
2:00 271.4 F 28.2
3:00 272.6 F 29.2
4:00 271.9 F 28.5

Min 271.3 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.2 F 0.0
5:00 212.3 F 0.0
10:00 233.8 F 0.0
15:00 232.5 F 0.0
20:00 222.7 F 0.0
25:00 224.9 F 0.0
30:00 225.0 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 20:48
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:02
CYCLE COMPLETE

04/28/2022 02:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____