
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001603
15:58 11/19/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 0:39
STOP BUTTON PRESSED
NOT STERILE
END TIME 0:39

Digital Signature #
AE29BBF4ABF733CF
