

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 003647  
10:52 11/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:12  
PRESSURIZING 2:16  
STERILIZING 3:14  
136.4 C 223kPa 3:14  
136.3 C 223kPa 3:44  
136.3 C 224kPa 4:14  
136.4 C 224kPa 4:44  
136.4 C 225kPa 5:14  
136.3 C 223kPa 5:44  
136.4 C 225kPa 6:14  
136.4 C 223kPa 6:44  
136.4 C 223kPa 6:44  
VENTING 6:44  
AIR DRYING 7:36  
CYCLE COMPLETE 7:49

-----  
Digital Signature #  
69F59266A4526C0E  
-----