
Midmark M11 - v1.0.5
Total Cycles: 3277

Name: RITTER2
Sterilizer ID: V2313228
01/04/2024 01:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.1 F 0.0
2:00 158.4 F 2.3
4:00 180.2 F 5.9
6:00 200.8 F 10.2
8:00 219.4 F 6.6
10:00 226.7 F 8.4
12:00 236.7 F 13.4
14:00 239.8 F 11.1
16:00 238.6 F 10.1
18:00 252.2 F 16.8
20:00 263.7 F 23.6
22:00 271.2 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.7
1:00 272.5 F 29.2
2:00 271.2 F 28.5
3:00 272.4 F 29.5
4:00 271.9 F 29.0

Min 271.0 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.0 F 0.0
5:00 199.5 F 0.0
10:00 199.8 F 0.0
15:00 198.0 F 0.0
20:00 193.2 F 0.0
25:00 188.7 F 0.0
30:00 185.1 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 22:02
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:00:54
CYCLE COMPLETE

01/04/2024 02:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____