

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2705

Name: RITTER2  
Sterilizer ID: V2313228  
07/26/2023 12:36 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 115.9 F 0.0  
2:00 128.4 F 0.0  
4:00 138.6 F 0.0  
6:00 174.0 F 2.3  
8:00 192.7 F 6.2  
10:00 218.0 F 10.2  
12:00 224.6 F 7.5  
14:00 233.9 F 9.2  
16:00 242.1 F 11.5  
18:00 241.0 F 11.6  
20:00 239.2 F 10.3  
22:00 252.1 F 16.6  
24:00 263.2 F 23.1  
26:00 272.3 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.6  
1:00 272.4 F 29.1  
2:00 271.5 F 28.8  
3:00 272.3 F 29.1  
4:00 271.4 F 28.6

Min 271.3 F 28.5  
Max 272.5 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.6 F 0.0  
5:00 217.8 F 0.0  
10:00 225.0 F 0.0  
15:00 222.0 F 0.0  
20:00 219.9 F 0.0  
25:00 222.4 F 0.0  
30:00 225.1 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 26:13  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:04:22  
CYCLE COMPLETE

07/26/2023 01:40 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_