

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2092

Name: RITTER 1  
Sterilizer ID: V2178198  
07/06/2022 07:29 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	73.5 F	0.0
2:00	78.0 F	0.0
4:00	83.1 F	0.0
6:00	111.9 F	0.0
8:00	154.9 F	0.0
10:00	181.5 F	0.0
12:00	207.0 F	0.0
14:00	212.1 F	0.0
16:00	212.6 F	0.1
18:00	212.6 F	0.1
20:00	212.7 F	0.1
22:00	212.7 F	0.1
24:00	212.7 F	0.1
26:00	212.6 F	0.0
28:00	212.5 F	0.0

C633: HEATUP MODE  
PRESSURE LOW  
CHAMBER TEMP 213F  
CHAMBER PRESS 0.0PSI  
\*\* ITEMS NOT STERILE \*\*

07/06/2022 07:59 AM