

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001337  
15:43 09/11/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:56  
PRESSURIZING 3:58  
STERILIZING 5:08  
136.0 C 220kPa 5:08  
136.3 C 224kPa 5:38  
136.4 C 225kPa 6:08  
136.4 C 225kPa 6:38  
136.4 C 225kPa 7:08  
136.2 C 223kPa 7:38  
136.3 C 224kPa 8:08  
136.3 C 223kPa 8:38  
136.3 C 223kPa 8:38  
VENTING 8:38  
AIR DRYING 9:34  
CYCLE COMPLETE 10:33  
-----

Digital Signature #  
05718136C7481950  
-----