
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.7uS / 0.4ppm
CYCLE NUMBER 000781
12:18 08/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:35
PRESSURIZING 6:20
STERILIZING 7:46
136.0 C 220kPa 7:46
136.1 C 221kPa 8:16
136.4 C 223kPa 8:46
136.4 C 218kPa 9:16
135.6 C 217kPa 9:46
136.1 C 220kPa 10:16
136.3 C 222kPa 10:46
136.4 C 223kPa 11:16
136.4 C 223kPa 11:16
VENTING 11:16
AIR DRYING 12:14
CYCLE COMPLETE 13:05

Digital Signature #
84D6B34D10E7BD47
