
Midmark M11 - v1.0.5
Total Cycles: 3715

Name: RITTER2
Sterilizer ID: V2313228
06/19/2024 08:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 70.6 F 0.0
4:00 82.2 F 0.3
6:00 105.2 F 1.8
8:00 138.0 F 4.5
10:00 160.8 F 7.6
12:00 178.7 F 10.9
14:00 204.7 F 5.4
16:00 221.8 F 6.7
18:00 224.6 F 10.5
20:00 227.6 F 8.4
22:00 237.0 F 10.6
24:00 234.6 F 8.3
26:00 245.6 F 13.7
28:00 256.2 F 19.4
30:00 266.1 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.2
1:00 271.8 F 29.7
2:00 272.3 F 29.6
3:00 271.4 F 29.5
4:00 272.3 F 29.8

Min 271.2 F 29.1
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.9 F 0.0
5:00 192.4 F 0.0
10:00 179.4 F 0.0
15:00 178.9 F 0.0
20:00 176.5 F 0.0
25:00 175.9 F 0.0
30:00 176.0 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:41
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:09:51
CYCLE COMPLETE

06/19/2024 09:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____