
Midmark M11 - v1.0.0
Total Cycles: 2653

Name: RITTER 1
Sterilizer ID: V2178198
12/07/2022 01:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 216.4 F 0.0
2:00 211.9 F 4.3
4:00 207.6 F 9.4
6:00 221.4 F 6.7
8:00 234.9 F 10.5
10:00 234.4 F 7.3
12:00 242.6 F 11.5
14:00 246.3 F 13.2
16:00 259.3 F 20.2
18:00 271.0 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.0
1:00 272.3 F 28.8
2:00 271.5 F 28.0
3:00 272.3 F 28.8
4:00 271.8 F 28.2

Min 271.2 F 27.9
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 242.0 F 0.0
10:00 238.6 F 0.0
15:00 228.1 F 0.0
20:00 226.3 F 0.0
25:00 228.5 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 18:34
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:57:24
CYCLE COMPLETE

12/07/2022 02:29 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____