
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001148
13:26 29/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:04
PRESSURIZING 4:05
STERILIZING 5:18
136.0 C 221kPa 5:18
136.4 C 223kPa 5:48
136.4 C 222kPa 6:18
136.3 C 224kPa 6:48
136.3 C 223kPa 7:18
136.3 C 224kPa 7:48
136.4 C 224kPa 8:18
136.4 C 224kPa 8:48
136.4 C 224kPa 8:48
VENTING 8:48
AIR DRYING 9:46
CYCLE COMPLETE 10:46

Digital Signature #
2D29533A8F059E87
