
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000678
15:11 14/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:09
PRESSURIZING 2:13
CYCLE ABORTED 2:16
STOP BUTTON PRESSED
NOT STERILE
END TIME 2:16

Digital Signature #
5843A3B25A2A3CBD
