
Midmark M11 - v1.0.5
Total Cycles: 3867

Name: RITTER2
Sterilizer ID: V2313228
08/16/2024 10:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.9 F 0.0
2:00 133.7 F 1.3
4:00 142.4 F 3.6
6:00 168.0 F 6.7
8:00 193.5 F 10.3
10:00 216.1 F 5.6
12:00 217.1 F 7.5
14:00 231.5 F 11.8
16:00 231.5 F 8.5
18:00 240.8 F 12.2
20:00 243.4 F 12.5
22:00 254.3 F 18.4
24:00 264.3 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.6
1:00 271.8 F 29.9
2:00 271.9 F 29.8
3:00 271.6 F 29.6
4:00 271.5 F 29.5

Min 271.3 F 29.4
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.9 F 0.0
5:00 228.9 F 0.0
10:00 191.2 F 0.0
15:00 187.6 F 0.0
20:00 184.7 F 0.0
25:00 185.7 F 0.0
30:00 184.6 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 25:57
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:04:51
CYCLE COMPLETE

08/16/2024 12:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____