

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1050

Name: RITTER2  
Sterilizer ID: V2313228  
04/28/2022 03:06 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 200.3 F 0.0  
2:00 197.5 F 3.9  
4:00 206.1 F 9.4  
6:00 221.2 F 7.7  
8:00 232.4 F 11.1  
10:00 234.2 F 9.1  
12:00 238.6 F 9.7  
14:00 248.7 F 15.2  
16:00 261.1 F 22.3  
18:00 272.4 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.2  
1:00 272.8 F 29.9  
2:00 272.0 F 29.1  
3:00 272.3 F 29.9  
4:00 272.6 F 29.6  
  
Min 271.6 F 29.0  
Max 273.0 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.8 F 0.0  
5:00 203.7 F 0.0  
10:00 221.7 F 0.0  
15:00 222.9 F 0.0  
20:00 220.0 F 0.0  
25:00 217.4 F 0.0  
30:00 215.3 F 0.0  
Drying Complete

FILLING: 2:16  
HEATING: 18:29  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:57:58  
CYCLE COMPLETE

04/28/2022 04:03 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_