
Midmark M11 - v1.0.5
Total Cycles: 738

Name: RITTER2
Sterilizer ID: V2313228
01/31/2022 03:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.8 F 0.0
2:00 129.2 F 1.8
4:00 153.9 F 5.1
6:00 183.2 F 9.7
8:00 206.2 F 5.9
10:00 223.3 F 7.8
12:00 235.9 F 13.3
14:00 238.7 F 11.1
16:00 238.7 F 10.1
18:00 252.9 F 17.1
20:00 264.8 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.7
1:00 272.5 F 29.5
2:00 271.4 F 28.6
3:00 272.1 F 29.6
4:00 272.4 F 29.2

Min 271.1 F 28.6
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.4 F 0.0
5:00 228.5 F 0.0
10:00 205.5 F 0.0
15:00 195.1 F 0.0
20:00 199.1 F 0.0
25:00 223.8 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 21:50
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:00:02
CYCLE COMPLETE

01/31/2022 04:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____