

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5420

Name: RITTER 1  
Sterilizer ID: V2178198  
03/17/2025 01:31 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 90.0 F 0.0  
2:00 91.9 F 0.9  
4:00 120.4 F 3.5  
6:00 154.5 F 6.9  
8:00 180.6 F 10.8  
10:00 195.5 F 3.1  
12:00 216.6 F 5.3  
14:00 230.0 F 9.2  
16:00 236.8 F 11.9  
18:00 238.1 F 9.8  
20:00 236.9 F 8.7  
22:00 247.6 F 13.8  
24:00 259.6 F 20.3  
26:00 270.5 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.1  
1:00 272.4 F 28.4  
2:00 272.3 F 28.7  
3:00 271.7 F 28.1  
4:00 272.7 F 28.8

Min 271.5 F 27.9  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.4 F 0.0  
5:00 199.7 F 0.0  
10:00 191.5 F 0.0  
15:00 194.2 F 0.0  
20:00 200.5 F 0.0  
25:00 209.5 F 0.0  
30:00 238.9 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 26:48  
STERILIZING: 4:00  
VENTING: 3:30  
DRYING: 30:00  
TOTAL CYCLE: 01:05:06  
CYCLE COMPLETE

03/17/2025 02:36 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_