
Midmark M11 - v1.0.0
Total Cycles: 4765

Name: RITTER 1
Sterilizer ID: V2178198
07/25/2024 11:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 112.0 F 0.0
2:00 121.4 F 0.7
4:00 124.9 F 2.9
6:00 145.4 F 5.7
8:00 166.7 F 9.0
10:00 190.2 F 12.8
12:00 208.5 F 4.6
14:00 220.0 F 8.4
16:00 232.1 F 12.6
18:00 231.8 F 8.6
20:00 243.0 F 12.9
22:00 241.6 F 11.5
24:00 253.1 F 17.2
26:00 263.4 F 23.3
28:00 272.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.6
1:00 272.0 F 29.4
2:00 272.4 F 29.2
3:00 272.4 F 29.3
4:00 271.9 F 29.4

Min 271.0 F 28.5
Max 273.1 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 200.6 F 0.0
10:00 199.3 F 0.0
15:00 216.8 F 0.0
20:00 242.6 F 0.0
25:00 234.4 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 28:23
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:06:59
CYCLE COMPLETE

07/25/2024 12:19 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____