
Midmark M11 - v1.0.0
Total Cycles: 2260

Name: RITTER 1
Sterilizer ID: V2178198
08/29/2022 02:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.9 F 0.0
2:00 176.5 F 0.5
4:00 198.1 F 1.7
6:00 212.6 F 6.2
8:00 227.9 F 11.0
10:00 230.2 F 7.3
12:00 243.0 F 13.2
14:00 241.7 F 11.5
16:00 240.2 F 10.5
18:00 254.4 F 17.6
20:00 266.3 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.4 F 28.8
2:00 271.7 F 28.5
3:00 272.3 F 28.8
4:00 271.5 F 28.4

Min 271.5 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 227.3 F 0.0
10:00 224.3 F 0.0
15:00 230.8 F 0.0
20:00 230.1 F 0.0
25:00 225.7 F 0.0
Dry Cycle Aborted

FILLING: 1:05
HEATING: 21:33
STERILIZING: 4:00
VENTING: 3:13
DRYING: 25:43
TOTAL CYCLE: 00:55:34

08/29/2022 03:39 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____