

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2672

Name: RITTER2  
Sterilizer ID: V2313228  
07/20/2023 07:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.4 F 0.0  
2:00 72.9 F 0.0  
4:00 82.5 F 0.0  
6:00 102.0 F 0.0  
8:00 138.5 F 0.2  
10:00 173.7 F 0.7  
12:00 189.3 F 4.1  
14:00 210.1 F 7.4  
16:00 220.7 F 10.8  
18:00 226.4 F 7.6  
20:00 231.7 F 10.0  
22:00 235.5 F 9.4  
24:00 242.6 F 12.4  
26:00 240.5 F 10.9  
28:00 252.6 F 17.0  
30:00 263.3 F 23.3  
32:00 272.7 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.8  
1:00 272.5 F 29.6  
2:00 271.7 F 28.7  
3:00 272.2 F 29.6  
4:00 272.3 F 29.2  
  
Min 271.4 F 28.6  
Max 272.8 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.0 F 0.0  
5:00 199.8 F 0.0  
10:00 211.9 F 0.0  
15:00 211.6 F 0.0  
20:00 211.7 F 0.0  
25:00 212.0 F 0.0  
30:00 210.8 F 0.0  
  
Drying Complete

FILLING: 0:48  
HEATING: 32:19  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:10:15  
CYCLE COMPLETE

07/20/2023 08:48 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_