
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000114
13:10 03/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:38
PRESSURIZING 4:38
STERILIZING 6:06
136.5 C 226kPa 6:06
136.4 C 226kPa 6:36
136.4 C 226kPa 7:06
136.2 C 225kPa 7:36
136.3 C 225kPa 8:06
136.4 C 223kPa 8:36
136.4 C 226kPa 9:06
136.2 C 224kPa 9:36
136.2 C 224kPa 9:36
VENTING 9:36
AIR DRYING 10:36
CYCLE COMPLETE 25:36

Digital Signature #
4D47A29BA7C21A39
