
Midmark M11 - v1.0.5
Total Cycles: 732

Name: RITTER2
Sterilizer ID: V2313228
01/26/2022 12:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.1 F 0.0
2:00 151.3 F 2.8
4:00 178.4 F 7.0
6:00 200.9 F 12.0
8:00 217.1 F 6.4
10:00 230.9 F 11.8
12:00 233.7 F 9.2
14:00 238.5 F 9.9
16:00 247.3 F 14.3
18:00 259.5 F 21.1
20:00 270.6 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.7 F 29.7
2:00 271.7 F 28.9
3:00 272.3 F 29.8
4:00 272.5 F 29.4

Min 271.4 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.9 F 0.0
5:00 201.9 F 0.0
10:00 212.1 F 0.0
15:00 218.9 F 0.0
20:00 221.8 F 0.0
25:00 221.6 F 0.0
30:00 223.6 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 20:49
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:32
CYCLE COMPLETE

01/26/2022 01:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____