

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001405  
12:14 09/08/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:34  
PRESSURIZING 6:05  
STERILIZING 7:23  
136.0 C 220kPa 7:23  
135.9 C 219kPa 7:53  
136.2 C 221kPa 8:23  
136.4 C 223kPa 8:53  
136.2 C 221kPa 9:23  
136.4 C 223kPa 9:53  
136.2 C 222kPa 10:23  
136.4 C 221kPa 10:53  
VENTING 10:53  
AIR DRYING 11:56  
CYCLE COMPLETE 11:58  
-----

Digital Signature #  
E575F8E8500C91EC  
-----