
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007591
7:53 25/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:17
PRESSURIZING 1:26
STERILIZING 2:23
136.0 C 220kPa 2:23
136.4 C 222kPa 2:53
136.4 C 219kPa 3:23
136.1 C 219kPa 3:53
136.4 C 220kPa 4:23
136.2 C 220kPa 4:53
136.4 C 223kPa 5:23
136.4 C 222kPa 5:53
136.4 C 222kPa 5:53
VENTING 5:53
AIR DRYING 6:43
CYCLE COMPLETE 8:46

Digital Signature #
042A7F8074DA783A
