
Midmark M11 - v1.0.5
Total Cycles: 3481

Name: RITTER2
Sterilizer ID: V2313228
03/27/2024 11:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.2 F 0.0
2:00 146.1 F 1.0
4:00 151.9 F 3.7
6:00 178.4 F 7.7
8:00 196.2 F 12.1
10:00 217.3 F 6.2
12:00 223.4 F 10.9
14:00 230.5 F 7.6
16:00 243.6 F 13.1
18:00 243.1 F 12.6
20:00 255.4 F 18.9
22:00 266.4 F 25.7
24:00 273.1 F 30.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.7
1:00 272.3 F 29.7
2:00 272.9 F 30.1
3:00 272.0 F 30.2
4:00 272.0 F 29.7

Min 271.6 F 29.6
Max 272.9 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.7 F 0.0
5:00 223.1 F 0.0
10:00 228.8 F 0.0
15:00 206.5 F 0.0
20:00 209.2 F 0.0
25:00 215.4 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 24:19
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:03:01
CYCLE COMPLETE

03/27/2024 12:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____