
Midmark M11 - v1.0.5
Total Cycles: 248

Name: RITTER2
Sterilizer ID: V2313228
07/26/2021 01:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.6 F 0.0
2:00 195.5 F 4.6
4:00 215.4 F 11.7
6:00 227.1 F 9.1
8:00 234.5 F 9.1
10:00 239.6 F 10.5
12:00 249.2 F 15.3
14:00 262.3 F 22.8
16:00 272.1 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 271.7 F 29.1
2:00 272.3 F 29.5
3:00 272.0 F 29.7
4:00 271.7 F 29.1

Min 271.4 F 28.9
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.2 F 0.0
5:00 228.5 F 0.0
10:00 235.3 F 0.0
15:00 229.5 F 0.0
20:00 203.5 F 0.0
25:00 195.1 F 0.0
30:00 185.0 F 0.0
Drying Complete

FILLING: 2:39
HEATING: 16:08
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 00:55:46
CYCLE COMPLETE

07/27/2021 06:19 AM
ACCEPTED: X
REJECTED:
Operator: U01