
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001186
9:49 22/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:01
PRESSURIZING 3:13
STERILIZING 4:40
136.0 C 221kPa 4:40
136.2 C 223kPa 5:10
136.4 C 224kPa 5:40
136.4 C 222kPa 6:10
136.3 C 223kPa 6:40
136.4 C 225kPa 7:10
136.4 C 225kPa 7:40
136.2 C 223kPa 8:10
136.2 C 223kPa 8:10
VENTING 8:10
AIR DRYING 9:02
CYCLE COMPLETE 24:02

Digital Signature #
F3A54B70CC4453DD
