

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1816

Name: RITTER 1  
Sterilizer ID: V2178198  
04/13/2022 07:29 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.2 F 0.0  
2:00 77.1 F 0.0  
4:00 90.4 F 0.0  
6:00 126.9 F 0.1  
8:00 159.0 F 0.3  
10:00 176.7 F 3.3  
12:00 198.6 F 6.6  
14:00 208.6 F 9.9  
16:00 219.0 F 13.4  
18:00 224.5 F 6.9  
20:00 234.9 F 11.0  
22:00 233.2 F 7.2  
24:00 243.8 F 12.8  
26:00 241.4 F 10.9  
28:00 253.6 F 17.1  
30:00 264.3 F 23.4  
32:00 271.6 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.5 F 28.9  
2:00 271.4 F 28.2  
3:00 272.5 F 29.0  
4:00 272.0 F 28.4  
  
Min 271.3 F 28.1  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.1 F 0.0  
5:00 212.3 F 0.0  
10:00 206.0 F 0.0  
15:00 196.0 F 0.0  
20:00 206.6 F 0.0  
25:00 229.6 F 0.0  
30:00 228.9 F 0.0  
  
Drying Complete

FILLING: 0:45  
HEATING: 32:07  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:09:60  
CYCLE COMPLETE

04/13/2022 08:38 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_