
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006759
8:26 19/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:42
PRESSURIZING 3:42
STERILIZING 4:44
136.4 C 220kPa 4:44
136.4 C 221kPa 5:14
135.9 C 220kPa 5:44
136.4 C 221kPa 6:14
136.3 C 220kPa 6:44
136.3 C 220kPa 7:14
136.4 C 223kPa 7:44
136.3 C 222kPa 8:14
136.3 C 222kPa 8:14
VENTING 8:14
AIR DRYING 9:05
CYCLE COMPLETE 11:44

Digital Signature #
33B890B8B6802225
