
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000478
13:50 17/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:43
PRESSURIZING 1:47
STERILIZING 2:26
136.0 C 220kPa 2:26
136.3 C 222kPa 2:56
136.4 C 222kPa 3:26
136.4 C 224kPa 3:56
136.0 C 222kPa 4:26
136.2 C 222kPa 4:56
136.3 C 224kPa 5:26
136.4 C 224kPa 5:56
136.4 C 224kPa 5:56
VENTING 5:57
CYCLE ABORTED 6:23
STOP BUTTON PRESSED
NOT STERILE
END TIME 6:23

Digital Signature #
F515A39261D09018
