
Midmark M11 - v1.0.0
Total Cycles: 1287

Name: RITTER 1
Sterilizer ID: V2178198
10/22/2021 06:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.4 F 0.0
2:00 72.8 F 0.0
4:00 78.3 F 0.0
6:00 110.0 F 0.0
8:00 140.6 F 0.1
10:00 173.0 F 0.2
12:00 194.9 F 0.5
14:00 209.1 F 3.7
16:00 223.3 F 6.5
18:00 232.2 F 9.7
20:00 235.1 F 13.4
22:00 237.9 F 9.8
24:00 238.4 F 9.4
26:00 245.6 F 13.0
28:00 242.9 F 11.5
30:00 255.1 F 17.8
32:00 265.9 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.4 F 28.8
2:00 271.3 F 28.1
3:00 272.5 F 28.8
4:00 271.3 F 28.0

Min 271.2 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.9 F 0.0
5:00 218.7 F 0.0
10:00 198.3 F 0.0
15:00 187.4 F 0.0
20:00 181.1 F 0.0
25:00 177.2 F 0.0
30:00 173.3 F 0.0

Drying Complete

FILLING: 0:54
HEATING: 33:47
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:11:46
CYCLE COMPLETE

10/22/2021 08:08 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____