
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002444
13:04 23/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:02
PRESSURIZING 2:06
STERILIZING 3:05
136.0 C 222kPa 3:05
136.4 C 226kPa 3:35
136.2 C 224kPa 4:05
136.2 C 221kPa 4:35
136.0 C 223kPa 5:05
136.4 C 224kPa 5:35
136.4 C 223kPa 6:05
136.4 C 226kPa 6:35
136.4 C 226kPa 6:35
VENTING 6:35
AIR DRYING 7:25
CYCLE COMPLETE 37:25

Digital Signature #
D51F2620A8C904A1
