

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002211  
11:03 10/08/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:33  
PRESSURIZING 3:34  
STERILIZING 4:46  
136.4 C 220kPa 4:46  
136.1 C 222kPa 5:16  
136.5 C 225kPa 5:46  
136.4 C 222kPa 6:16  
136.4 C 225kPa 6:46  
135.9 C 221kPa 7:16  
136.4 C 224kPa 7:46  
136.4 C 225kPa 8:16  
136.4 C 225kPa 8:16  
VENTING 8:16  
AIR DRYING 9:12  
CYCLE COMPLETE 9:13

-----  
Digital Signature #  
800C01047B90EAA6  
-----