

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000226  
9:19 28/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:58  
PRESSURIZING 3:59  
STERILIZING 5:21  
136.0 C 228kPa 5:21  
136.2 C 231kPa 5:51  
136.4 C 232kPa 6:21  
136.3 C 233kPa 6:51  
136.4 C 233kPa 7:21  
136.3 C 233kPa 7:51  
136.4 C 233kPa 8:21  
136.4 C 232kPa 8:51  
136.4 C 232kPa 8:51  
VENTING 8:51  
AIR DRYING 9:52  
CYCLE COMPLETE 18:55  
-----

Digital Signature #  
C67CA581A64A0BDF  
-----