
Midmark M11 - v1.0.5
Total Cycles: 1005

Name: RITTER2
Sterilizer ID: V2313228
04/19/2022 11:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.5 F 0.0
2:00 153.8 F 0.3
4:00 177.8 F 1.7
6:00 199.4 F 6.0
8:00 214.7 F 10.7
10:00 224.8 F 6.4
12:00 236.3 F 12.3
14:00 237.3 F 10.5
16:00 236.9 F 9.6
18:00 251.4 F 16.6
20:00 263.1 F 23.5
22:00 272.3 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.5 F 29.8
2:00 271.5 F 29.0
3:00 272.1 F 29.9
4:00 272.1 F 29.4

Min 271.2 F 28.9
Max 272.5 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.9 F 0.0
5:00 205.5 F 0.0
10:00 207.5 F 0.0
15:00 209.4 F 0.0
20:00 215.6 F 0.0
25:00 221.7 F 0.0
30:00 224.2 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 22:12
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:52
CYCLE COMPLETE

04/19/2022 12:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____