
Midmark M11 - v1.0.5
Total Cycles: 823

Name: RITTER2
Sterilizer ID: V2313228
02/23/2022 07:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.8 F 0.0
2:00 75.9 F 0.0
4:00 110.1 F 0.1
6:00 117.9 F 0.3
8:00 154.2 F 3.4
10:00 178.1 F 7.1
12:00 199.5 F 11.0
14:00 213.6 F 4.8
16:00 224.5 F 8.8
18:00 235.0 F 13.4
20:00 236.3 F 10.3
22:00 235.7 F 8.6
24:00 247.9 F 14.7
26:00 259.5 F 21.2
28:00 270.0 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.4 F 29.5
2:00 271.6 F 29.2
3:00 272.4 F 29.4
4:00 271.7 F 29.3

Min 271.4 F 28.9
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.7 F 0.0
5:00 199.7 F 0.0
10:00 208.0 F 0.0
15:00 210.9 F 0.0
20:00 209.8 F 0.0
25:00 213.6 F 0.0
30:00 222.5 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 28:56
STERILIZING: 4:00
VENTING: 2:56
DRYING: 30:00
TOTAL CYCLE: 01:06:48
CYCLE COMPLETE

02/23/2022 09:04 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____