
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000047
8:50 08/09/2020
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:45
PRESSURIZING 8:21
STERILIZING 10:24
136.2 C 225kPa 10:24
136.5 C 226kPa 10:54
136.4 C 227kPa 11:24
136.4 C 224kPa 11:54
136.3 C 225kPa 12:24
136.4 C 226kPa 12:55
136.4 C 227kPa 13:25
136.4 C 227kPa 13:54
VENTING 13:54
AIR DRYING 15:09
CYCLE COMPLETE 45:09

Digital Signature #
85416873DE888889
