
Midmark M11 - v1.0.5
Total Cycles: 1642

Name: RITTER2
Sterilizer ID: V2313228
10/24/2022 08:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.9 F 0.0
2:00 181.2 F 2.7
4:00 204.8 F 6.4
6:00 221.0 F 11.3
8:00 219.6 F 6.1
10:00 227.1 F 11.8
12:00 233.8 F 8.7
14:00 239.3 F 10.5
16:00 245.9 F 13.3
18:00 257.8 F 19.8
20:00 268.6 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.7 F 29.4
2:00 271.8 F 28.8
3:00 271.4 F 29.2
4:00 272.8 F 29.3

Min 271.1 F 28.6
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.6 F 0.0
5:00 222.7 F 0.0
10:00 231.5 F 0.0
15:00 225.7 F 0.0
20:00 222.0 F 0.0
25:00 221.8 F 0.0
30:00 220.2 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 21:13
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:40
CYCLE COMPLETE

10/24/2022 09:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____