
Midmark M11 - v1.0.5
Total Cycles: 1511

Name: RITTER2
Sterilizer ID: V2313228
09/21/2022 07:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.1 F 0.0
2:00 78.8 F 0.0
4:00 96.9 F 0.0
6:00 149.4 F 0.1
8:00 175.9 F 1.6
10:00 208.5 F 5.2
12:00 215.4 F 9.1
14:00 226.5 F 13.1
16:00 226.7 F 7.2
18:00 236.2 F 11.6
20:00 234.2 F 8.5
22:00 242.2 F 11.3
24:00 245.0 F 13.2
26:00 257.1 F 19.7
28:00 268.0 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.6 F 29.7
2:00 271.5 F 28.9
3:00 272.7 F 29.9
4:00 271.9 F 29.2

Min 271.2 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.4 F 0.0
5:00 204.0 F 0.0
10:00 220.4 F 0.0
15:00 233.6 F 0.0
20:00 230.6 F 0.0
25:00 227.6 F 0.0
30:00 226.8 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 29:19
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:07:13
CYCLE COMPLETE

09/21/2022 08:36 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____