
Midmark M11 - v1.0.5
Total Cycles: 491

Name: RITTER2
Sterilizer ID: V2313228
10/27/2021 11:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.3 F 0.0
2:00 183.8 F 4.4
4:00 197.2 F 10.3
6:00 219.8 F 6.4
8:00 233.0 F 13.0
10:00 235.2 F 11.6
12:00 240.3 F 10.8
14:00 253.8 F 17.9
16:00 265.4 F 25.4
18:00 272.5 F 30.9

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 30.4
1:00 273.0 F 30.4
2:00 272.6 F 30.7
3:00 272.5 F 30.9
4:00 272.4 F 30.2

Min 272.1 F 30.2
Max 273.0 F 31.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.6 F 0.0
5:00 229.9 F 0.0
10:00 227.2 F 0.0
15:00 229.7 F 0.0
20:00 205.3 F 0.0
25:00 205.4 F 0.0
30:00 197.9 F 0.0
Drying Complete

FILLING: 2:18
HEATING: 18:25
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 00:57:44
CYCLE COMPLETE

10/27/2021 01:04 PM
ACCEPTED: X
REJECTED:
Operator: U01